

## NOTES:


- 1. SUBSTRATE: N-BK7
- 2. S2 TO BE PARALLEL WITH \$1 TO <1 ARCMIN
- 3. COATING (APPLY ACROSS CLEAR APERTURE) S1 & S2: RABS ≤0.25% @ 880nm RAVG ≤1.25% @ 400 - 870 nm RAVG ≤1.25% @ 890 - 1000nm



- 5. Rohs Compliant
- 6. SURFACE FLATNESS (P-V):  $\lambda/4$  (per inch over clear aperture)







## **PARTS TO THIS DRAWING**

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

	\$1	\$2
SHAPE	PLANO	PLANO
SURFACE QUALITY	60-40	60-40
CLEAR APERTURE	Ø11.25	Ø11.25
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED

			Edmund Optic	;S®
THIRD ANGLE PROJECTION	$\phi$	TITLE	12.5mm DIA., 3mm THICK, VIS-NIR CO λ/4 N-BK7 WINDOW	ATED
ALL DIMS IN	mm	DWG NO	45658	SHEET 1 OF

12.5mm DIA., 3mm THICK, VIS-NIR COATED
λ/4 N-BK7 WINDOW

L DIMS IN	mm	DWG NO	45658	SHEE 1 OF