NOTES:

1. SUBSTRATE MATERIAL:

N-BK7

2. COATING (APPLY ACROSS CLEAR APERTURE)

S1 & S2: VIS-NIR

R(abs) ≤0.25% @ 880nm


R(avg) ≤1.25% @ 400 - 870nm


R(avg) ≤1.25% @ 890 - 1000nm

3. FINE GROUND SURFACE

4. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICAIONS APPLY ACROSS CLEAR APERTURE

5. Rohs Compliant





FOR INFORMATION ONLY: PARTS TO THIS DRAWING

> SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

REV. A	\$1	S2					
SHAPE	CONVEX	PLANO	EFL	50		Edmund Ontice	OR
RADIUS	25.8	8	BFL	49.34		Edmund Optic	5 5°
SURFACE QUALITY	60-40	60-40	THIRD ANGLE PROJECTION		TITLE	5mm H X 10mm L X 50mm FL VIS-NIR COATED, CYLINDER LENS	
CLEAR APERTURE	9.00 X 4.00	9.00 X 4.00					
COATING APERTURE	9.00 X 4.00	9.00 X 4.00					
BEVEL	PROTECTED AS NEEDED	PROTECTED AS NEEDED	ALL DIMS IN	mm	DWG NO	47960	SHEET 1 OF 1