
## NOTES:

- 1. SUBSTRATE: N-BK7
- 2. S2 TO BE PARALLEL TO S1 TO WITHIN <3 ARCMINS
- 3. COATING (APPLY ACROSS COATING APERTURE) \$1 & \$2: R(ABS) < 0.25% @ 355nm @ 0° AOI

DAMAGE THRESHOLD PULSED: 7.5J/cm² @ 20ns, 20Hz @ 355nm

- 4. FINE GROUND SURFACE
- 5. CLEAR APERTURE AND COATING APERTURE ARE CENTERED ON SURFACE
- 6. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 7. ROHS COMPLIANT



## FOR INFORMATION ONLY: DO NOT MANUFACTURE PARTS TO THIS DRAWING

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

SHAPE	S1 PLANO	S2 PLANO		8	<b>Edmund Option</b>	CS®
SURFACE QUALITY	10-5	10-5	1	TITLE	Ø25.4mm x 6.35mm 355nm N-BK7 LASER WINDOW	
SURFACE FLATNESS	λ/10	λ/10	THIRD ANGLE PROJECTION			
MIN CLEAR APERTURE	Ø22.86	Ø22.86			711112311	CHEET
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN mm	DWG NO	38933	SHEET 1 OF 1