- SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)


\$1: HARD DIELECTRIC \$PUTTERED T(avg): ≥91% FROM 408 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 390nm @ 0° AOI T(abs): =50% FOR 400±4nm @ 0° AOI


S2:SINGLE LAYER MgF2


4.\FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT









LONGPASS FILTER

REV A	\$1	\$2
SHAPE	PLANO	PLANO
SURFACE QUALITY	40-20	40-20
CLEAR APERTURE	>80%	>80%
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED



			•	
THIRD ANGLE - PROJECTION		TITLE	Ø25mm, 400nm, HIGH PERFORMANG LONGPASS FILTER	CE
ALL DIMS IN	mm	DWG NO	62981	SHEET 1 OF 1

- SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)


\$1: HARD DIELECTRIC \$PUTTERED T(avg): ≥91% FROM 458 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 440nm @ 0° AOI T(abs): =50% FOR 450±4.5nm @ 0° AOI


S2:SINGLE LAYER MgF2


4.\FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT









LONGPASS FILTER

REV A	\$1	\$2
SHAPE	PLANO	PLANO
SURFACE QUALITY	40-20	40-20
CLEAR APERTURE	>80%	>80%
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED



THIRD ANGLE _ PROJECTION	ϕ	TITLE	Ø25mm, 450nm, HIGH PERFORMAN LONGPASS FILTER	CE
ALL DIMS IN	mm	DWG NO	62982	SHEET 1 OF 1

- SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)


\$1: HARD DIELECTRIC \$PUTTERED T(avg): ≥91% FROM 508 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 490nm @ 0° AOI T(abs): =50% FOR 500±5nm @ 0° AOI


S2:SINGLE LAYER MgF2


4.\FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT









LONGPASS FILTER

REV A	\$1	\$2	
SHAPE	PLANO	PLANO	
SURFACE QUALITY	40-20	40-20	
CLEAR APERTURE	>80%	>80%	
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	



THIRD ANGLE PROJECTION	ϕ	TITLE	Ø25mm, 500nm, HIGH PERFORMANC LONGPASS FILTER	
ALL DIMS IN	mm	DWG NO	62983 SF	HEET OF 1

- SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)


\$1: HARD DIELECTRIC \$PUTTERED T(avg): ≥91% FROM 560 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 539nm @ 0° AOI T(abs): =50% FOR 550±5.5nm @ 0° AOI


S2:SINGLE LAYER MgF2


4.\FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT









LONGPASS FILTER

REV A	\$1	\$2
SHAPE	PLANO	PLANO
SURFACE QUALITY	40-20	40-20
CLEAR APERTURE	>80%	>80%
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

Edmund Optics®

)		
THIRD ANGLE PROJECTION		TITLE	Ø25mm, 550nm, HIGH P LONGPASS FII	CE
ALL DIMS IN	mm	DWG NO	62984	SHEET 1 OF 1

- SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)


\$1: HARD DIELECTRIC \$PUTTERED T(avg): ≥91% FROM 610 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 588nm @ 0° AOI T(abs): =50% FOR 600±6nm @ 0° AOI


S2:SINGLE LAYER MgF2


4.\FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT









LONGPASS FILTER

REV A	\$1	\$2
SHAPE	PLANO	PLANO
SURFACE QUALITY	40-20	40-20
CLEAR APERTURE	>80%	>80%
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED



			<u> </u>	
THIRD ANGLE _ PROJECTION	ϕ	TITLE	Ø25mm, 600nm, HIGH PERFO LONGPASS FILTER	RMANCE
ALL DIMS IN	mm	DWG NO	62985	SHEET 1 OF 1

- SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)


\$1: HARD DIELECTRIC \$PUTTERED T(avg): ≥91% FROM 660 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 637nm @ 0° AOI T(abs): =50% FOR 650±6.5nm @ 0° AOI


S2:SINGLE LAYER MgF2


4.\FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT









LONGPASS FILTER

REV A	\$1	\$2	
SHAPE	PLANO	PLANO	
SURFACE QUALITY	40-20	40-20	
CLEAR APERTURE	>80%	>80%	
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	



THIRD ANGLE - PROJECTION	ϕ	TITLE	Ø25mm, 650nm, HIGH PERFORMAN LONGPASS FILTER	CE
ALL DIMS IN	mm	DWG NO	62986	SHEET 1 OF 1

- SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)


\$1: HARD DIELECTRIC \$PUTTERED T(avg): ≥91% FROM 710 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 686nm @ 0° AOI T(abs): =50% FOR 700±7nm @ 0° AOI


S2:SINGLE LAYER MgF2


4.\FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT









LONGPASS FILTER

REV A	\$1	\$2			
SHAPE	PLANO	ANO PLANO			
SURFACE QUALITY	40-20	40-20			
CLEAR APERTURE	>80%	>80%			
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED			



				•	
THIRD ANGLE TITLE		TITLE	Ø25mm, 700nm, HIGH PERFORMANCE LONGPASS FILTER		
ALL DIMS IN	mm	DWG NO	62987	SHEET 1 OF 1	